

Demotek™

EXAMINATION GLOVES

Demotek - Medical Related Products by Demophorius Limited. Designed and produced for all medical use. The quality assurance of these products is particularly high complying with all international requirements and manufactured under strict GMP standards.



PRODUCT INFORMATION

Name of Product: LATEX EXAMINATION GLOVES

Type : Pre-Powdered or Powder-Free - Non-Sterilized

Material : Natural high grade latex

Dimension Table :

Glove Size	Palm Width (mm)	230 max
Extra-Small	75±10	230 max
Small	80±10	230 max
Medium	95±10	230 max
Large	110±10	230 max
Extra Large	115±10	230 max

Size	Shortest Length	Cuff	Palm	Finger	Unaged	aged 100 C22HR
S	230 max	0.08 max	0.08 max	0.08 max	18±3 MPa	16±2 MPa
M	230 max	0.08 max	0.08 max	0.08 max	18±3 MPa	16±2 MPa
L	230 max	0.08 max	0.08 max	0.08 max	18±3 MPa	16±2 MPa

Test Reports :

	PH	Water insoluble salt	Chloride	Sulphate	Protein
Glove Inside	7.15±0.2	0.4±0.1mg/g	≥ 0.1	≥ 0.2	≥ 50.0
Glove Outside	7.15±0.2	0.8±0.1mg/g	Ïg/c m2	Ïg/c m2	Ïg/c m2

Powder : FDA approved USP absorbable cornstarch

Hypoallergenic : Yes (laboratory test reports available on request)

Colour : Natural

Design and Feature : Ambidextrous, straight fingers, smooth or textured surface and beaded cuff

Packing : 100 pieces of gloves per dispenser box, 20 dispenser boxes per carton

Pre Shipment Quality Inspection Table :

Criteria	Insp. Level	AQL
Dimension	S-2	4.0
Physical Properties	S-2	4.0
1000ml Water Leak Test	G-I	1.5
Visual Inspection for major defects	G-I	2.5
Visual Inspection for minor defects	G-I	4.0



BS EN ISO 9002:1994



BS EN ISO 9002:1994



EN 46001 (1993) AND
EN 29001 (1987)



COUNCIL DIRECTIVE
93/42 EEC

Distributed by:

Product Conformance : In compliance with ASTM D3578-95 and BS EN 455 (Part 1&2)

Quality Assurance : Manufactured under Good Manufacturing Practice (GMP) and BS EN ISO 9002 Quality Systems

This Product information is subject to change without prior notice.



Demophorus
Limited

196 Archbishop Makarios III Ave., CY-3030 Limassol, Cyprus

Tel.: ++ 357 5 749896, Fax: ++ 357 5 749907, E-mail: demotek@spidernet.com.cy